

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025270.D
 Acq On : 23 Nov 2021 13:34
 Operator : JC/MD
 Sample : M4788-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-B-320-341-111921

Quant Time: Nov 23 15:21:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	99413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	177988	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58253	55.591	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.180%	
35) Dibromofluoromethane	5.391	113	57253	52.426	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.860%	
50) Toluene-d8	8.653	98	207595	50.558	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.120%	
62) 4-Bromofluorobenzene	11.079	95	77653	49.990	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.980%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.319	96	6325	6.865	ug/l	83
16) Acetone	2.380	43	25962	44.588	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	7578	7.507	ug/l	85
24) 1,1-Dichloroethane	3.611	63	5672	3.269	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	460925	391.026	ug/l	81
42) 1,2-Dichloroethane	6.092	62	2788	1.821	ug/l	97
44) Trichloroethene	7.129	130	174945	128.020	ug/l	94
52) Toluene	8.720	92	26513	8.633	ug/l	99
64) Tetrachloroethene	9.275	164	1300	0.937	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025270.D
Acq On : 23 Nov 2021 13:34
Operator : JC/MD
Sample : M4788-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-B-320-341-111921

Quant Time: Nov 23 15:21:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

