

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025272.D
 Acq On : 23 Nov 2021 14:21
 Operator : JC/MD
 Sample : M4801-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20211119

Quant Time: Nov 23 15:22:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58112	55.049	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 110.100%		
35) Dibromofluoromethane	5.391	113	57156	52.615	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 105.240%		
50) Toluene-d8	8.653	98	208867	51.138	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 102.280%		
62) 4-Bromofluorobenzene	11.085	95	77213	49.971	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 99.940%		
Target Compounds						Qvalue
11) Tert butyl alcohol	2.965	59	7419	20.193	ug/l	99
16) Acetone	2.380	43	8571	14.612	ug/l	90
25) 2-Butanone	4.568	43	2522	2.974	ug/l	94
68) m/p-Xylenes	10.299	106	1001	0.415	ug/l	90
69) o-Xylene	10.646	106	741	0.314	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025272.D
Acq On : 23 Nov 2021 14:21
Operator : JC/MD
Sample : M4801-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20211119

Quant Time: Nov 23 15:22:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

