

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025276.D
 Acq On : 23 Nov 2021 15:54
 Operator : JC/MD
 Sample : M4800-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW9-GW-788-790

Quant Time: Nov 24 03:13:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	168967	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57077	56.782	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.560%	
35) Dibromofluoromethane	5.391	113	53902	51.993	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.980%	
50) Toluene-d8	8.653	98	198880	51.022	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.040%	
62) 4-Bromofluorobenzene	11.085	95	72765	49.344	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.680%	
Target Compounds						
16) Acetone	2.380	43	4018	7.194	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025276.D
Acq On : 23 Nov 2021 15:54
Operator : JC/MD
Sample : M4800-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW9-GW-788-790

Quant Time: Nov 24 03:13:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

