

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025284.D
 Acq On : 23 Nov 2021 19:01
 Operator : JC/MD
 Sample : M4809-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1994

Quant Time: Nov 24 03:15:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	94493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57594	57.824	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	115.640%	
35) Dibromofluoromethane	5.391	113	54780	52.031	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.060%	
50) Toluene-d8	8.653	98	199826	50.480	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.960%	
62) 4-Bromofluorobenzene	11.079	95	77432	51.706	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	103.420%	
Target Compounds						
16) Acetone	2.380	43	3066	5.540	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025284.D
Acq On : 23 Nov 2021 19:01
Operator : JC/MD
Sample : M4809-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1994

Quant Time: Nov 24 03:15:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

