

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025287.D
 Acq On : 23 Nov 2021 20:08
 Operator : JC/MD
 Sample : M4803-06 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 261-262

Quant Time: Nov 24 03:15:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.574	168	60401	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	6.775	114	76548	50.000	ug/l	0.01
63) Chlorobenzene-d5	10.055	117	40277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	11213	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	9476	14.884	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	29.760%#
35) Dibromofluoromethane	5.397	113	13737	29.248	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	58.500%#
50) Toluene-d8	8.653	98	80811	45.762	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.520%#
62) 4-Bromofluorobenzene	11.085	95	13388	20.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	40.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025287.D
Acq On : 23 Nov 2021 20:08
Operator : JC/MD
Sample : M4803-06 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
261-262

Quant Time: Nov 24 03:15:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

