

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025297.D
 Acq On : 24 Nov 2021 00:24
 Operator : JC/MD
 Sample : M4758-03
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP2

Quant Time: Nov 24 03:18:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	170097	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	163975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56331	55.638	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.280%	
35) Dibromofluoromethane	5.397	113	53682	51.436	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.880%	
50) Toluene-d8	8.653	98	203817	51.941	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.880%	
62) 4-Bromofluorobenzene	11.085	95	77155	51.974	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.940%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025297.D
Acq On : 24 Nov 2021 00:24
Operator : JC/MD
Sample : M4758-03
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP2

Quant Time: Nov 24 03:18:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

