

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025298.D
 Acq On : 24 Nov 2021 00:48
 Operator : JC/MD
 Sample : M4758-05
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3

Quant Time: Nov 24 03:18:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	94226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	168429	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	160055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73071	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55912	56.294	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.580%
35) Dibromofluoromethane	5.397	113	53106	51.388	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.780%
50) Toluene-d8	8.653	98	199950	51.460	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.920%
62) 4-Bromofluorobenzene	11.085	95	74148	50.443	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025298.D
Acq On : 24 Nov 2021 00:48
Operator : JC/MD
Sample : M4758-05
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3

Quant Time: Nov 24 03:18:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

