

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025300.D
 Acq On : 24 Nov 2021 01:34
 Operator : JC/MD
 Sample : M4758-15
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP8

Quant Time: Nov 24 03:18:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	95076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	163251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57255	57.131	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.260%
35) Dibromofluoromethane	5.397	113	54096	51.279	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.560%
50) Toluene-d8	8.653	98	202986	51.176	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.360%
62) 4-Bromofluorobenzene	11.085	95	76464	50.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.920%
Target Compounds						
11) Tert butyl alcohol	2.953	59	8111m	24.660	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025300.D
Acq On : 24 Nov 2021 01:34
Operator : JC/MD
Sample : M4758-15
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP8

Quant Time: Nov 24 03:18:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

