

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025301.D
 Acq On : 24 Nov 2021 01:58
 Operator : JC/MD
 Sample : M4758-21
 Misc : 5.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP11

Quant Time: Nov 24 03:18:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	97451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	170775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57943	56.409	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.820%
35) Dibromofluoromethane	5.397	113	55707	51.241	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.480%
50) Toluene-d8	8.653	98	207721	50.818	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.640%
62) 4-Bromofluorobenzene	11.085	95	79637	51.499	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.000%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.947	59	6444	17.029	ug/l #	73
20) Methylene Chloride	2.782	84	1314	1.225	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025301.D
Acq On : 24 Nov 2021 01:58
Operator : JC/MD
Sample : M4758-21
Misc : 5.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP11

Quant Time: Nov 24 03:18:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

