

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025303.D
 Acq On : 24 Nov 2021 02:45
 Operator : JC/MD
 Sample : M4803-03
 Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 263

Quant Time: Nov 24 03:18:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	98905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178830	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	167367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	78322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58908	56.505	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.020%
35) Dibromofluoromethane	5.397	113	55211	50.318	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.640%
50) Toluene-d8	8.653	98	211599	51.291	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.580%
62) 4-Bromofluorobenzene	11.085	95	77577	49.706	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.420%
Target Compounds						
18) Methyl Acetate	2.727	43	7759	4.192	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025303.D
Acq On : 24 Nov 2021 02:45
Operator : JC/MD
Sample : M4803-03
Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
263

Quant Time: Nov 24 03:18:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

