

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025307.D
 Acq On : 24 Nov 2021 04:18
 Operator : JC/MD
 Sample : M4795-07
 Misc : 4.72g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HAR-25

Quant Time: Nov 24 05:02:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193312	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	184815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63412	55.540	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.080%	
35) Dibromofluoromethane	5.397	113	60242	50.790	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.580%	
50) Toluene-d8	8.653	98	228989	51.348	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.085	95	86814	51.457	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.920%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025307.D
Acq On : 24 Nov 2021 04:18
Operator : JC/MD
Sample : M4795-07
Misc : 4.72g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HAR-25

Quant Time: Nov 24 05:02:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

