

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025311.D
 Acq On : 24 Nov 2021 05:51
 Operator : JC/MD
 Sample : M4812-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SETTING-TANK-01

Quant Time: Nov 24 09:56:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107899	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201126	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	65612	57.690	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.380%
35) Dibromofluoromethane	5.391	113	63303	52.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.100%
50) Toluene-d8	8.653	98	239582	52.892	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.780%
62) 4-Bromofluorobenzene	11.085	95	92497	53.978	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.960%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.136	74	4274	5.488	ug/l	94
16) Acetone	2.380	43	14956	23.666	ug/l	98
20) Methylene Chloride	2.788	84	4754	4.002	ug/l	95
43) Isopropyl Acetate	6.348	43	9470	2.920	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025311.D
Acq On : 24 Nov 2021 05:51
Operator : JC/MD
Sample : M4812-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SETTING-TANK-01

Quant Time: Nov 24 09:56:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

