



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025289.D
 Acq On : 23 Nov 2021 21:18
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Title : SW846 8260
 Last Update : Wed Nov 17 15:36:16 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	4490	PASS
75	95	30	60	47.4	12659	PASS
95	95	100	100	100.0	26688	PASS
96	95	5	9	6.8	1811	PASS
173	174	0.00	2	0.6	146	PASS
174	95	50	100	89.9	23987	PASS
175	174	5	9	7.3	1746	PASS
176	174	95	101	97.0	23272	PASS
177	176	5	9	6.4	1484	PASS