

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033014.D
 Acq On : 23 Nov 2022 13:07
 Operator : JC/MD
 Sample : N5730-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20221117DL

Quant Time: Nov 24 04:05:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	170043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	148481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60166	58.983	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	117.960%#
35) Dibromofluoromethane	5.385	113	47087	49.119	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	8.647	98	160019	55.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	111.460%
62) 4-Bromofluorobenzene	11.079	95	57272	45.480	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.960%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	875	0.886	ug/l	93
44) Trichloroethene	7.129	130	96593	80.096	ug/l	99
64) Tetrachloroethene	9.275	164	889	0.846	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033014.D
Acq On : 23 Nov 2022 13:07
Operator : JC/MD
Sample : N5730-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20221117DL

Quant Time: Nov 24 04:05:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

