

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033016.D
 Acq On : 23 Nov 2022 13:53
 Operator : JC/MD
 Sample : N5730-07DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D6-20221118DL

Quant Time: Nov 24 04:05:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211373	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73577	57.664	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.320%
35) Dibromofluoromethane	5.391	113	58298	48.922	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.840%
50) Toluene-d8	8.647	98	202605	56.801	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	113.600%#
62) 4-Bromofluorobenzene	11.079	95	72185	46.114	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.220%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	686	0.556	ug/l	86
44) Trichloroethene	7.129	130	43734	29.174	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033016.D
Acq On : 23 Nov 2022 13:53
Operator : JC/MD
Sample : N5730-07DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D6-20221118DL

Quant Time: Nov 24 04:05:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

