

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033017.D
 Acq On : 23 Nov 2022 14:16
 Operator : JC/MD
 Sample : N5729-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20221117RE

Quant Time: Nov 24 04:06:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72996	59.349	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 118.700%#		
35) Dibromofluoromethane	5.385	113	57102	48.982	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.960%		
50) Toluene-d8	8.647	98	197173	56.493	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 112.980%#		
62) 4-Bromofluorobenzene	11.079	95	69313	45.262	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.520%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	415	0.427	ug/l	74
9) 1,1,2-Trichlorotrifluo...	2.331	101	5293	4.444	ug/l	99
12) 1,1-Dichloroethene	2.325	96	665	0.599	ug/l	83
27) cis-1,2-Dichloroethene	4.513	96	1436	0.996	ug/l	83
44) Trichloroethene	7.129	130	87734	59.824	ug/l	97
52) Toluene	8.720	92	2214	0.644	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033017.D
Acq On : 23 Nov 2022 14:16
Operator : JC/MD
Sample : N5729-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20221117RE

Quant Time: Nov 24 04:06:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

