

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033019.D
 Acq On : 23 Nov 2022 15:02
 Operator : JC/MD
 Sample : N5741-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Nov 24 04:06:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201332	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70211	58.187	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.380%
35) Dibromofluoromethane	5.391	113	54791	48.273	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.540%
50) Toluene-d8	8.647	98	191692	56.407	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	112.820%#
62) 4-Bromofluorobenzene	11.079	95	67212	45.079	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.160%
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	24329	6.452	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033019.D
Acq On : 23 Nov 2022 15:02
Operator : JC/MD
Sample : N5741-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Nov 24 04:06:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

