

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033024.D
 Acq On : 23 Nov 2022 16:57
 Operator : JC/MD
 Sample : N5730-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP06-20221118

Quant Time: Nov 24 04:07:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	207739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85429	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67932	55.178	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.360%
35) Dibromofluoromethane	5.385	113	54072	46.170	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.340%
50) Toluene-d8	8.647	98	188031	53.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.020%
62) 4-Bromofluorobenzene	11.079	95	66156	43.002	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033024.D
Acq On : 23 Nov 2022 16:57
Operator : JC/MD
Sample : N5730-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP06-20221118

Quant Time: Nov 24 04:07:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

