

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033025.D
 Acq On : 23 Nov 2022 17:20
 Operator : JC/MD
 Sample : N5744-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Nov 24 04:07:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199970	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66403	55.960	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.920%
35) Dibromofluoromethane	5.391	113	53126	47.124	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.240%
50) Toluene-d8	8.647	98	180485	53.354	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.700%
62) 4-Bromofluorobenzene	11.079	95	64902	43.826	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.660%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033025.D
Acq On : 23 Nov 2022 17:20
Operator : JC/MD
Sample : N5744-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Nov 24 04:07:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

