

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033026.D
 Acq On : 23 Nov 2022 17:43
 Operator : JC/MD
 Sample : N5744-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 24 04:07:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85165	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67168	55.337	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.680%
35) Dibromofluoromethane	5.385	113	53917	46.856	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.720%
50) Toluene-d8	8.647	98	185976	53.884	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.760%
62) 4-Bromofluorobenzene	11.079	95	64855	42.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033026.D
Acq On : 23 Nov 2022 17:43
Operator : JC/MD
Sample : N5744-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 24 04:07:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

