

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033031.D
 Acq On : 23 Nov 2022 19:39
 Operator : JC/MD
 Sample : N5741-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-30

Quant Time: Nov 24 04:08:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52707	50.968	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.940%
35) Dibromofluoromethane	5.385	113	45092	45.433	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.860%
50) Toluene-d8	8.653	98	134763	44.909	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.820%#
62) 4-Bromofluorobenzene	11.079	95	62671	48.070	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.140%
Target Compounds						Qvalue
16) Acetone	2.392	43	1829	3.795	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	3463	1.080	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033031.D
Acq On : 23 Nov 2022 19:39
Operator : JC/MD
Sample : N5741-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-30

Quant Time: Nov 24 04:08:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

