

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033038.D
 Acq On : 24 Nov 2022 04:55
 Operator : JC/MD
 Sample : N5741-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Nov 24 05:20:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	227796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	371437	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	105149	47.054	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.100%
35) Dibromofluoromethane	5.391	113	88531	42.278	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.560%
50) Toluene-d8	8.647	98	271721	42.817	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	85.640%#
62) 4-Bromofluorobenzene	11.079	95	122695	44.605	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.200%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.001	59	22292	63.391	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	46314	6.718	ug/l #	1
22) Diisopropyl ether	3.764	45	5369	0.768	ug/l #	83
31) Cyclohexane	5.477	56	96428	26.827	ug/l	92
39) Methylcyclohexane	7.379	83	16056	4.203	ug/l	91
40) Benzene	6.038	78	274277	29.216	ug/l	98
52) Toluene	8.720	92	4745	0.768	ug/l	97
67) Ethyl Benzene	10.195	91	165308	14.045	ug/l	100
68) m/p-Xylenes	10.305	106	3404	0.739	ug/l	98
69) o-Xylene	10.646	106	1033	0.232	ug/l	95
73) Isopropylbenzene	10.963	105	66469	5.894	ug/l	100
78) n-propylbenzene	11.305	91	70544	5.471	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	6076	0.643	ug/l	100
85) sec-Butylbenzene	11.890	105	8249	0.731	ug/l	99
89) n-Butylbenzene	12.329	91	3974	0.483	ug/l #	70
95) Naphthalene	13.774	128	53377	5.290	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033038.D
Acq On : 24 Nov 2022 04:55
Operator : JC/MD
Sample : N5741-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Nov 24 05:20:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

