

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033046.D
 Acq On : 24 Nov 2022 07:59
 Operator : JC/MD
 Sample : N5741-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-40

Quant Time: Nov 25 06:38:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	212701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70553	56.085	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 112.160%			
35) Dibromofluoromethane	5.385	113	56861	47.419	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.840%			
50) Toluene-d8	8.647	98	193169	53.699	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 107.400%			
62) 4-Bromofluorobenzene	11.079	95	70314	44.639	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.280%			
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.995	59	97062	485.067	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	53483	13.633	ug/l #	90
22) Diisopropyl ether	3.757	45	27596	6.940	ug/l #	86
39) Methylcyclohexane	7.385	83	1506	0.688	ug/l	97
40) Benzene	6.037	78	672903	125.172	ug/l	100
52) Toluene	8.720	92	2547	0.720	ug/l	93
67) Ethyl Benzene	10.195	91	5998	0.894	ug/l	94
68) m/p-Xylenes	10.305	106	1809	0.688	ug/l	99
69) o-Xylene	10.640	106	2669	1.050	ug/l	95
73) Isopropylbenzene	10.963	105	6092	0.957	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	3362	0.629	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	6060	1.137	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033046.D
Acq On : 24 Nov 2022 07:59
Operator : JC/MD
Sample : N5741-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-40

Quant Time: Nov 25 06:38:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

