

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033048.D
 Acq On : 24 Nov 2022 08:45
 Operator : JC/MD
 Sample : N5742-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-42

Quant Time: Nov 25 06:38:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69487	56.711	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.420%
35) Dibromofluoromethane	5.385	113	55113	47.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	188730	54.126	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.260%
62) 4-Bromofluorobenzene	11.079	95	68772	45.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.060%
Target Compounds						
16) Acetone	2.392	43	1561	2.717	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033048.D
Acq On : 24 Nov 2022 08:45
Operator : JC/MD
Sample : N5742-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-42

Quant Time: Nov 25 06:38:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

