

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033050.D
 Acq On : 24 Nov 2022 09:30
 Operator : JC/MD
 Sample : N5742-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-44

Quant Time: Nov 25 06:39:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210049	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69164	55.820	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.640%
35) Dibromofluoromethane	5.385	113	54145	45.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	8.647	98	187593	52.770	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.540%
62) 4-Bromofluorobenzene	11.079	95	67623	43.472	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.940%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.995	59	20519	104.137	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	22318	5.778	ug/l	97
22) Diisopropyl ether	3.764	45	6236	1.593	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033050.D
Acq On : 24 Nov 2022 09:30
Operator : JC/MD
Sample : N5742-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-44

Quant Time: Nov 25 06:39:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

