

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033009.D
 Acq On : 23 Nov 2022 11:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 24 04:04:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	118	0.00
2 T	Dichlorodifluoromethane	0.382	0.350	8.4	112	0.00
3 P	Chloromethane	0.412	0.367	10.9	111	0.00
4 C	Vinyl Chloride	0.459	0.427	7.0#	113	0.00
5 T	Bromomethane	0.340	0.293	13.8	109	0.01
6 T	Chloroethane	0.374	0.347	7.2	108	0.00
7 T	Trichlorofluoromethane	0.891	0.867	2.7	118	0.00
8 T	Diethyl Ether	0.281	0.281	0.0	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.473	-1.1	119	0.00
10 T	Methyl Iodide	0.536	0.439	18.1	90	0.00
11 T	Tert butyl alcohol	0.077	0.088	-14.3	135	0.06
12 CM	1,1-Dichloroethene	0.436	0.429	1.6#	116	0.00
13 T	Acrolein	0.089	0.080	10.1	111	0.00
14 T	Allyl chloride	0.704	0.712	-1.1	118	0.00
15 T	Acrylonitrile	0.218	0.223	-2.3	123	0.00
16 T	Acetone	0.227	0.258	-13.7	134	0.02
17 T	Carbon Disulfide	1.205	1.113	7.6	114	0.00
18 T	Methyl Acetate	0.629	0.644	-2.4	124	0.00
19 T	Methyl tert-butyl Ether	1.513	1.537	-1.6	118	0.00
20 T	Methylene Chloride	0.548	0.477	13.0	115	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.471	5.6	113	0.00
22 T	Diisopropyl ether	1.534	1.581	-3.1	116	0.00
23 T	Vinyl Acetate	1.201	1.255	-4.5	117	0.00
24 P	1,1-Dichloroethane	0.878	0.865	1.5	117	0.00
25 T	2-Butanone	0.314	0.341	-8.6	127	0.00
26 T	2,2-Dichloropropane	0.706	0.743	-5.2	120	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.568	-0.2	115	0.00
28 T	Bromochloromethane	0.386	0.376	2.6	112	-0.01
29 T	Tetrahydrofuran	0.199	0.204	-2.5	122	0.01
30 C	Chloroform	0.983	0.977	0.6#	114	0.00
31 T	Cyclohexane	0.789	0.783	0.8	115	0.00
32 T	1,1,1-Trichloroethane	0.880	0.888	-0.9	115	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.477	11.5	115	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	118	0.00
35 S	Dibromofluoromethane	0.282	0.261	7.4	115	0.00
36 T	1,1-Dichloropropene	0.451	0.448	0.7	115	0.00
37 T	Ethyl Acetate	0.393	0.416	-5.9	121	0.00
38 T	Carbon Tetrachloride	0.500	0.506	-1.2	116	0.00
39 T	Methylcyclohexane	0.514	0.530	-3.1	118	0.00
40 TM	Benzene	1.264	1.262	0.2	115	0.00
41 T	Methacrylonitrile	0.221	0.237	-7.2	122	0.00
42 TM	1,2-Dichloroethane	0.509	0.509	0.0	114	0.00
43 T	Isopropyl Acetate	0.650	0.692	-6.5	121	0.00
44 TM	Trichloroethene	0.355	0.355	0.0	113	0.00
45 C	1,2-Dichloropropane	0.312	0.319	-2.2#	115	0.00
46 T	Dibromomethane	0.244	0.242	0.8	114	0.00
47 T	Bromodichloromethane	0.471	0.490	-4.0	116	0.00
48 T	Methyl methacrylate	0.335	0.351	-4.8	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033009.D
 Acq On : 23 Nov 2022 11:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 24 04:04:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	133	0.04
50 S	Toluene-d8	0.915	0.832	9.1	117	0.00
51 T	4-Methyl-2-Pentanone	0.401	0.434	-8.2	122	0.00
52 CM	Toluene	0.831	0.851	-2.4#	115	0.00
53 T	t-1,3-Dichloropropene	0.462	0.520	-12.6	118	0.00
54 T	cis-1,3-Dichloropropene	0.508	0.542	-6.7	117	0.00
55 T	1,1,2-Trichloroethane	0.338	0.341	-0.9	116	0.00
56 T	Ethyl methacrylate	0.466	0.509	-9.2	117	0.00
57 T	1,3-Dichloropropane	0.546	0.561	-2.7	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.254	-14.9	130	0.00
59 T	2-Hexanone	0.296	0.332	-12.2	123	0.00
60 T	Dibromochloromethane	0.364	0.389	-6.9	115	0.00
61 T	1,2-Dibromoethane	0.357	0.371	-3.9	116	0.00
62 S	4-Bromofluorobenzene	0.370	0.358	3.2	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	119	0.00
64 T	Tetrachloroethene	0.354	0.328	7.3	111	0.00
65 PM	Chlorobenzene	0.997	0.984	1.3	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.381	-4.4	118	0.00
67 C	Ethyl Benzene	1.768	1.817	-2.8#	115	0.00
68 T	m/p-Xylenes	0.692	0.712	-2.9	116	0.00
69 T	o-Xylene	0.670	0.686	-2.4	115	0.00
70 T	Styrene	1.096	1.157	-5.6	114	0.00
71 P	Bromoform	0.280	0.305	-8.9	119	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
73 T	Isopropylbenzene	3.573	3.568	0.1	116	0.00
74 T	N-amyl acetate	1.215	1.292	-6.3	117	0.00
75 P	1,1,2,2-Tetrachloroethane	1.078	1.054	2.2	116	0.00
76 T	1,2,3-Trichloropropane	0.947	1.076	-13.6	133	0.00
77 T	Bromobenzene	0.879	0.840	4.4	112	0.00
78 T	n-propylbenzene	4.085	4.167	-2.0	116	0.00
79 T	2-Chlorotoluene	2.481	2.431	2.0	115	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.040	-1.3	114	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.323	-11.0	125	0.00
82 T	4-Chlorotoluene	2.908	2.882	0.9	114	0.00
83 T	tert-Butylbenzene	2.988	2.947	1.4	112	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.026	-1.1	114	0.00
85 T	sec-Butylbenzene	3.574	3.655	-2.3	115	0.00
86 T	p-Isopropyltoluene	3.028	3.127	-3.3	115	0.00
87 T	1,3-Dichlorobenzene	1.648	1.607	2.5	115	0.00
88 T	1,4-Dichlorobenzene	1.688	1.629	3.5	113	0.00
89 T	n-Butylbenzene	2.609	2.760	-5.8	118	0.00
90 T	Hexachloroethane	0.498	0.513	-3.0	118	0.00
91 T	1,2-Dichlorobenzene	1.603	1.565	2.4	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.226	-5.1	117	0.00
93 T	1,2,4-Trichlorobenzene	0.988	0.992	-0.4	116	0.00
94 T	Hexachlorobutadiene	0.436	0.430	1.4	121	0.00
95 T	Naphthalene	3.197	3.236	-1.2	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033009.D
Acq On : 23 Nov 2022 11:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 24 04:04:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.991	0.986	0.5	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6