

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
 Data File : VX033011.D
 Acq On : 23 Nov 2022 11:54
 Operator : JC/MD
 Sample : VX1123WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1123WBL01

Quant Time: Nov 24 04:04:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203245	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63329	51.338	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.680%	
35) Dibromofluoromethane	5.385	113	53215	46.443	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.880%	
50) Toluene-d8	8.647	98	161827	46.802	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 93.600%	
62) 4-Bromofluorobenzene	11.079	95	63966	42.498	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 85.000%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112322\
Data File : VX033011.D
Acq On : 23 Nov 2022 11:54
Operator : JC/MD
Sample : VX1123WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1123WBL01

Quant Time: Nov 24 04:04:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX033011.D
Acq: 23 Nov 2022 11:54




Instrument :
MSVOA_X
ClientSampleId :
VX1123WBL01


Tgt Ion:168 Resp: 126447
Ion Ratio Lower Upper
168 100
99 59.2 46.1 69.1



#33
1,2-Dichloroethane-d4
Concen: 51.338 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX033011.D
Acq: 23 Nov 2022 11:54

Tgt Ion: 65 Resp: 63329
Ion Ratio Lower Upper
65 100
67 50.4 0.0 101.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX033011.D

Acq: 23 Nov 2022 11:54

Instrument :

MSVOA_X

ClientSampleId :

VX1123WBL01





Tgt Ion:114 Resp: 203245

Ion Ratio Lower Upper

114 100

63 20.5 0.0 41.6

88 15.9 0.0 33.4



#35

Dibromofluoromethane

Concen: 46.443 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX033011.D

Acq: 23 Nov 2022 11:54




Tgt Ion:113 Resp: 53215


Ion Ratio Lower Upper

113 100

111 102.5 82.4 123.6

192 18.5 15.4 23.2









#50
Toluene-d8
Concen: 46.802 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX033011.D
Acq: 23 Nov 2022 11:54




Instrument :
MSVOA_X
ClientSampleId :
VX1123WBL01


Tgt Ion: 98 Resp: 161827
Ion Ratio Lower Upper
98 100
100 63.8 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 42.498 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX033011.D
Acq: 23 Nov 2022 11:54

Tgt Ion: 95 Resp: 63966
Ion Ratio Lower Upper
95 100
174 82.4 0.0 164.0
176 79.7 0.0 154.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX033011.D
Acq: 23 Nov 2022 11:54

Instrument :
MSVOA_X
ClientSampleId :
VX1123WBL01

Tgt Ion:117 Resp: 178829
Ion Ratio Lower Upper
117 100
82 53.9 44.3 66.5
119 32.1 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX033011.D
Acq: 23 Nov 2022 11:54

Tgt Ion:152 Resp: 81898
Ion Ratio Lower Upper
152 100
115 61.1 44.0 132.0
150 156.8 0.0 351.8

