

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019595.D
 Acq On : 24 Nov 2020 10:52
 Operator : JC/MD
 Sample : VX1124WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1124WBL01

Quant Time: Nov 25 04:39:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	313262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	511344	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	473916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199979	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	191648	47.18	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.36%		
35) Dibromofluoromethane	5.46	113	152319	47.46	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.92%		
50) Toluene-d8	8.69	98	617032	49.47	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.94%		
62) 4-Bromofluorobenzene	11.12	95	220825	49.21	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.42%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019595.D
Acq On : 24 Nov 2020 10:52
Operator : JC/MD
Sample : VX1124WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1124WBL01

Quant Time: Nov 25 04:39:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

