

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019598.D
 Acq On : 24 Nov 2020 12:25
 Operator : JC/MD
 Sample : L4839-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B1-GW

Quant Time: Nov 25 04:54:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	335898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	545394	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	491073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	201707	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	204213	46.89	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.78%
35) Dibromofluoromethane	5.46	113	163214	47.68	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.36%
50) Toluene-d8	8.70	98	657792	49.45	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.90%
62) 4-Bromofluorobenzene	11.12	95	221608	46.30	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.60%

Target Compounds

					Qvalue	
16) Acetone	2.42	43	5482	3.255	ug/l	93
17) Carbon Disulfide	2.54	76	2796	0.312	ug/l #	91
18) Methyl Acetate	2.75	43	1429	0.976	ug/l #	85
20) Methylene Chloride	2.83	84	1290	0.306	ug/l #	82
44) Trichloroethene	7.20	130	1294	0.314	ug/l	87
95) Naphthalene	13.82	128	2413	2.333	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019598.D
Acq On : 24 Nov 2020 12:25
Operator : JC/MD
Sample : L4839-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B1-GW

Quant Time: Nov 25 04:54:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

