

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019599.D
 Acq On : 24 Nov 2020 12:49
 Operator : JC/MD
 Sample : L4839-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B2-GW

Quant Time: Nov 25 10:21:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	321227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	528621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	474262	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	196194	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	198445	47.65	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.30%		
35) Dibromofluoromethane	5.46	113	156502	47.17	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.34%		
50) Toluene-d8	8.70	98	628081	48.71	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.42%		
62) 4-Bromofluorobenzene	11.12	95	218689	47.14	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.28%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	5692	3.534 ug/l	99
17) Carbon Disulfide	2.54	76	2107	0.246 ug/l #	87
20) Methylene Chloride	2.83	84	1208	0.300 ug/l #	90
44) Trichloroethene	7.18	130	1100	0.276 ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019599.D
Acq On : 24 Nov 2020 12:49
Operator : JC/MD
Sample : L4839-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B2-GW

Quant Time: Nov 25 10:21:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

