

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019600.D
 Acq On : 24 Nov 2020 13:12
 Operator : JC/MD
 Sample : L4841-02 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16855

Quant Time: Nov 25 10:23:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	314776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	522513	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199346	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	195649	47.94	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.88%		
35) Dibromofluoromethane	5.46	113	153608	46.83	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.66%		
50) Toluene-d8	8.70	98	627108	49.20	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.40%		
62) 4-Bromofluorobenzene	11.12	95	214230	46.72	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.44%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	39632	25.110	ug/l 99
20) Methylene Chloride	2.83	84	1059	0.268	ug/l 91
25) 2-Butanone	4.65	43	2823	1.191	ug/l 90
52) Toluene	8.76	92	7222	0.784	ug/l 100
68) m/p-Xylenes	10.34	106	4179	0.668	ug/l 98
69) o-Xylene	10.68	106	1967	0.329	ug/l 95
80) 1,3,5-Trimethylbenzene	11.49	105	3857	0.333	ug/l 100
84) 1,2,4-Trimethylbenzene	11.79	105	13411	1.146	ug/l 99
95) Naphthalene	13.82	128	20649	3.521	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019600.D
Acq On : 24 Nov 2020 13:12
Operator : JC/MD
Sample : L4841-02 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
16855

Quant Time: Nov 25 10:23:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

