

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019601.D
 Acq On : 24 Nov 2020 13:35
 Operator : JC/MD
 Sample : L4841-03 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16881

Quant Time: Nov 25 10:27:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	292542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	485696	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	445319	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	191872	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	184194	48.56	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.12%
35) Dibromofluoromethane	5.46	113	145025	47.57	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.14%
50) Toluene-d8	8.70	98	580729	49.02	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.04%
62) 4-Bromofluorobenzene	11.12	95	202592	47.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.06%

Target Compounds					Qvalue
16) Acetone	2.42	43	20993	14.312	ug/l 100
25) 2-Butanone	4.64	43	2362	1.072	ug/l 95
84) 1,2,4-Trimethylbenzene	11.79	105	3492	0.310	ug/l 89
95) Naphthalene	13.82	128	4670	2.494	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019601.D
Acq On : 24 Nov 2020 13:35
Operator : JC/MD
Sample : L4841-03 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
16881

Quant Time: Nov 25 10:27:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

