

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019606.D
 Acq On : 24 Nov 2020 15:32
 Operator : JC/MD
 Sample : L4839-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B5-GW

Quant Time: Nov 25 10:47:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	299585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	488780	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	427954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	163724	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	185854	47.85	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.70%	
35) Dibromofluoromethane		5.46	113	145988	47.58	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.16%	
50) Toluene-d8		8.70	98	580765	48.71	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.42%	
62) 4-Bromofluorobenzene		11.12	95	189540	44.19	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	4464	2.972	ug/l	90
18) Methyl Acetate	2.75	43	1131	0.938	ug/l	96
20) Methylene Chloride	2.84	84	1145	0.305	ug/l #	91
95) Naphthalene	13.82	128	2460	2.373	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019606.D
Acq On : 24 Nov 2020 15:32
Operator : JC/MD
Sample : L4839-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B5-GW

Quant Time: Nov 25 10:47:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

