

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019612.D
 Acq On : 24 Nov 2020 17:51
 Operator : JC/MD
 Sample : PB133213ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133213ZHE#06

Quant Time: Nov 25 05:47:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	293150	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	488358	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	432876	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	162960	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	183671	48.32	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.64%		
35) Dibromofluoromethane	5.46	113	145799	47.56	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.12%		
50) Toluene-d8	8.70	98	587149	49.29	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.58%		
62) 4-Bromofluorobenzene	11.12	95	187798	43.82	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	87.64%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	10607	7.216	ug/l 94
18) Methyl Acetate	2.75	43	4734	1.931	ug/l 98
20) Methylene Chloride	2.83	84	7928	2.158	ug/l 97
43) Isopropyl Acetate	6.42	43	76363	10.569	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019612.D
Acq On : 24 Nov 2020 17:51
Operator : JC/MD
Sample : PB133213ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133213ZHE#06

Quant Time: Nov 25 05:47:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

