

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019614.D
 Acq On : 24 Nov 2020 18:38
 Operator : JC/MD
 Sample : L4597-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BU-03-112320

Quant Time: Nov 25 05:55:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	324400	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	540466	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	491732	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	200050	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	203330	48.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.68%	
35) Dibromofluoromethane		5.46	113	161447	47.59	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.18%	
50) Toluene-d8		8.70	98	657258	49.86	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.72%	
62) 4-Bromofluorobenzene		11.12	95	224142	47.26	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	12427	7.640	ug/l	97
18) Methyl Acetate	2.75	43	4729	1.805	ug/l	98
20) Methylene Chloride	2.83	84	10295	2.532	ug/l	91
43) Isopropyl Acetate	6.41	43	76110	9.518	ug/l	100
95) Naphthalene	13.82	128	1153	2.253	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019614.D
Acq On : 24 Nov 2020 18:38
Operator : JC/MD
Sample : L4597-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BU-03-112320

Quant Time: Nov 25 05:55:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

