

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019621.D
 Acq On : 24 Nov 2020 22:07
 Operator : JC/MD
 Sample : VX1124WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1124WBL02

Quant Time: Nov 25 06:31:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	306207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	510723	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	464931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	193942	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	192506	48.49	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.98%		
35) Dibromofluoromethane	5.46	113	149921	46.77	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.54%		
50) Toluene-d8	8.70	98	615275	49.39	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.78%		
62) 4-Bromofluorobenzene	11.12	95	210145	46.89	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.78%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019621.D
Acq On : 24 Nov 2020 22:07
Operator : JC/MD
Sample : VX1124WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1124WBL02

Quant Time: Nov 25 06:31:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

