

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019625.D
 Acq On : 24 Nov 2020 23:40
 Operator : JC/MD
 Sample : L4842-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16846-41629

Quant Time: Nov 25 06:38:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	303188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	502686	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	439890	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	200252	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	190061	48.35	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.70%
35) Dibromofluoromethane	5.46	113	149056	47.24	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.48%
50) Toluene-d8	8.70	98	599735	48.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.82%
62) 4-Bromofluorobenzene	11.12	95	212136	48.09	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.18%

Target Compounds

					Qvalue
11) Tert butyl alcohol	3.04	59	12105	16.271	ug/l 96
16) Acetone	2.42	43	215218	141.570	ug/l 100
18) Methyl Acetate	2.75	43	12120	3.845	ug/l 96
20) Methylene Chloride	2.83	84	7004	1.843	ug/l 89
25) 2-Butanone	4.64	43	55613	24.353	ug/l 97
37) Ethyl Acetate	4.79	43	6217	1.405	ug/l 98
40) Benzene	6.11	78	679239	46.910	ug/l 100
44) Trichloroethene	7.18	130	1054	0.278	ug/l 69
51) 4-Methyl-2-Pentanone	8.62	43	4414	1.018	ug/l 96
52) Toluene	8.76	92	900830	101.697	ug/l 99
67) Ethyl Benzene	10.23	91	55722	3.517	ug/l 100
68) m/p-Xylenes	10.34	106	120547	20.422	ug/l 98
69) o-Xylene	10.68	106	119914	21.251	ug/l 99
78) n-propylbenzene	11.35	91	14877	0.936	ug/l # 69
80) 1,3,5-Trimethylbenzene	11.49	105	22243	1.914	ug/l 100
84) 1,2,4-Trimethylbenzene	11.79	105	100909	8.582	ug/l 93
95) Naphthalene	13.82	128	1326359	88.063	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019625.D
Acq On : 24 Nov 2020 23:40
Operator : JC/MD
Sample : L4842-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
16846-41629

Quant Time: Nov 25 06:38:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

