

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019628.D
 Acq On : 25 Nov 2020 00:50
 Operator : JC/MD
 Sample : L4835-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140B-20201119

Quant Time: Nov 25 06:46:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	310487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	517591	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	467320	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195255	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	192908	47.92	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.84%		
35) Dibromofluoromethane	5.46	113	151666	46.68	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.36%		
50) Toluene-d8	8.70	98	617829	48.94	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.88%		
62) 4-Bromofluorobenzene	11.12	95	219386	48.30	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.60%		

Target Compounds					Qvalue
16) Acetone	2.42	43	7236	4.648 ug/l	99
27) cis-1,2-Dichloroethene	4.56	96	22554	5.622 ug/l	93
44) Trichloroethene	7.18	130	49062	12.563 ug/l	99
95) Naphthalene	13.82	128	4420	2.472 ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019628.D
Acq On : 25 Nov 2020 00:50
Operator : JC/MD
Sample : L4835-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140B-20201119

Quant Time: Nov 25 06:46:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

