

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019629.D
 Acq On : 25 Nov 2020 01:13
 Operator : JC/MD
 Sample : L4835-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140N-20201119

Quant Time: Nov 25 06:47:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	327324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	536558	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	489468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	201232	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	199206	46.94	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.88%	
35) Dibromofluoromethane		5.46	113	157997	46.91	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.82%	
50) Toluene-d8		8.70	98	643348	49.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.32%	
62) 4-Bromofluorobenzene		11.12	95	224620	47.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	3305	2.014	ug/l #	88
27) cis-1,2-Dichloroethene	4.56	96	3062	0.724	ug/l	89
44) Trichloroethene	7.18	130	5708	1.410	ug/l	95
95) Naphthalene	13.82	128	2304	2.327	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019629.D
Acq On : 25 Nov 2020 01:13
Operator : JC/MD
Sample : L4835-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140N-20201119

Quant Time: Nov 25 06:47:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

