

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019634.D
 Acq On : 25 Nov 2020 03:09
 Operator : JC/MD
 Sample : L4835-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-118.5N-20201119

Quant Time: Nov 25 06:55:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	315736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	519909	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	464231	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	180279	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	199147	48.65	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.30%
35) Dibromofluoromethane	5.46	113	155975	47.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.58%
50) Toluene-d8	8.70	98	626550	49.41	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.82%
62) 4-Bromofluorobenzene	11.12	95	208444	45.69	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.38%

Target Compounds

					Qvalue
44) Trichloroethene	7.19	130	4063	1.036	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019634.D
Acq On : 25 Nov 2020 03:09
Operator : JC/MD
Sample : L4835-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-118.5N-20201119

Quant Time: Nov 25 06:55:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

