

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019635.D
 Acq On : 25 Nov 2020 03:32
 Operator : JC/MD
 Sample : L4835-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GR-OF-20201119

Quant Time: Nov 25 06:56:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	296962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	483878	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	437603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	174930	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	183707	47.71	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.42%
35) Dibromofluoromethane	5.46	113	145618	47.94	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.88%
50) Toluene-d8	8.70	98	585938	49.64	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.28%
62) 4-Bromofluorobenzene	11.12	95	202015	47.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.14%

Target Compounds					Qvalue
16) Acetone	2.42	43	3007	2.019 ug/l	97
44) Trichloroethene	7.19	130	3116	0.853 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019635.D
Acq On : 25 Nov 2020 03:32
Operator : JC/MD
Sample : L4835-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GR-OF-20201119

Quant Time: Nov 25 06:56:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

