

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019636.D
 Acq On : 25 Nov 2020 03:55
 Operator : JC/MD
 Sample : L4835-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPRING-20201119

Quant Time: Nov 25 07:03:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	317291	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	524061	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	191091	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	196607	47.79	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.58%
35) Dibromofluoromethane	5.46	113	155790	47.36	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.72%
50) Toluene-d8	8.70	98	626338	49.00	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.00%
62) 4-Bromofluorobenzene	11.12	95	211770	46.05	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.10%

Target Compounds

					Qvalue
16) Acetone	2.42	43	2838	1.784	ug/l 92
44) Trichloroethene	7.19	130	1658	0.419	ug/l 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019636.D
 Acq On : 25 Nov 2020 03:55
 Operator : JC/MD
 Sample : L4835-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPRING-20201119

Quant Time: Nov 25 07:03:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

