

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019637.D
 Acq On : 25 Nov 2020 04:18
 Operator : JC/MD
 Sample : L4835-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-118.5T-20201119

Quant Time: Nov 25 07:04:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	302870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	504734	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	454820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	179492	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	191808	48.84	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.68%
35) Dibromofluoromethane	5.46	113	149913	47.32	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.64%
50) Toluene-d8	8.70	98	605572	49.19	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.38%
62) 4-Bromofluorobenzene	11.12	95	205661	46.43	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.86%

Target Compounds

					Qvalue
16) Acetone	2.42	43	3835	2.525	ug/l 92
27) cis-1,2-Dichloroethene	4.56	96	13675	3.494	ug/l 91
30) Chloroform	5.17	83	10908	1.750	ug/l 99
44) Trichloroethene	7.18	130	6873	1.805	ug/l 87
47) Bromodichloromethane	7.88	83	3309	0.696	ug/l # 93
60) Dibromochloromethane	9.56	129	1617	0.456	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019637.D
Acq On : 25 Nov 2020 04:18
Operator : JC/MD
Sample : L4835-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-118.5T-20201119

Quant Time: Nov 25 07:04:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

