

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112420\
 Data File : VX019638.D
 Acq On : 25 Nov 2020 04:42
 Operator : JC/MD
 Sample : L4835-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121.5N-20201119

Quant Time: Nov 25 07:07:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	301070	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501865	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	453042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	181397	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	187054	47.92	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.84%
35) Dibromofluoromethane	5.46	113	147299	46.76	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.52%
50) Toluene-d8	8.70	98	601011	49.10	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.20%
62) 4-Bromofluorobenzene	11.12	95	203488	46.20	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.40%

Target Compounds					Qvalue
16) Acetone	2.42	43	2555	1.692 ug/l	98
44) Trichloroethene	7.18	130	2703	0.714 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112420\
Data File : VX019638.D
Acq On : 25 Nov 2020 04:42
Operator : JC/MD
Sample : L4835-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121.5N-20201119

Quant Time: Nov 25 07:07:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

