


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112420\
 Data File : VX019593.D
 Acq On : 24 Nov 2020 08:37
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Title : SW846 8260
 Last Update : Sat Nov 21 01:57:45 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	9497	PASS
75	95	30	60	47.8	27277	PASS
95	95	100	100	100.0	57122	PASS
96	95	5	9	6.5	3736	PASS
173	174	0.00	2	0.8	380	PASS
174	95	50	100	78.8	45010	PASS
175	174	5	9	7.7	3470	PASS
176	174	95	101	100.1	45056	PASS
177	176	5	9	6.2	2798	PASS