

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
 Data File : VX025320.D
 Acq On : 24 Nov 2021 12:42
 Operator : JC/MD
 Sample : M4775-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PK-10S-20211118

Quant Time: Nov 26 10:32:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189272	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63516	58.401	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.800%	
35) Dibromofluoromethane	5.385	113	61433	52.900	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.800%	
50) Toluene-d8	8.647	98	223790	51.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.500%	
62) 4-Bromofluorobenzene	11.079	95	84973	51.441	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.880%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
Data File : VX025320.D
Acq On : 24 Nov 2021 12:42
Operator : JC/MD
Sample : M4775-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PK-10S-20211118

Quant Time: Nov 26 10:32:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

