

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
 Data File : VX025328.D
 Acq On : 24 Nov 2021 15:49
 Operator : JC/MD
 Sample : M4775-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-12I-20211118RE

Quant Time: Nov 26 10:35:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	89933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165368	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56477	59.578	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 119.160%#	
35) Dibromofluoromethane	5.391	113	54654	53.865	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.740%	
50) Toluene-d8	8.653	98	197477	51.764	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.520%	
62) 4-Bromofluorobenzene	11.085	95	74811	51.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.680%	
Target Compounds						
					Qvalue	
24) 1,1-Dichloroethane	3.611	63	8289	5.280	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	5494	5.152	ug/l	84
40) Benzene	6.044	78	2680	0.613	ug/l	98
44) Trichloroethene	7.129	130	1807	1.423	ug/l	85
64) Tetrachloroethene	9.275	164	2195	1.637	ug/l #	85
65) Chlorobenzene	10.079	112	59756	18.032	ug/l	98
87) 1,3-Dichlorobenzene	11.975	146	5436	2.149	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	20612	8.041	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	11793m	4.783	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
Data File : VX025328.D
Acq On : 24 Nov 2021 15:49
Operator : JC/MD
Sample : M4775-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-12I-20211118RE

Quant Time: Nov 26 10:35:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

