

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
 Data File : VX025330.D
 Acq On : 24 Nov 2021 16:35
 Operator : JC/MD
 Sample : M4817-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A6879DL

Quant Time: Nov 26 10:36:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71243	50.245	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.500%			
35) Dibromofluoromethane	5.391	113	64360	48.883	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.760%			
50) Toluene-d8	8.653	98	247614	50.020	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.040%			
62) 4-Bromofluorobenzene	11.079	95	89289	47.678	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.360%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	41062	52.118	ug/l	95
25) 2-Butanone	4.562	43	61393	53.896	ug/l	96
29) Tetrahydrofuran	5.013	42	28492	40.397	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
Data File : VX025330.D
Acq On : 24 Nov 2021 16:35
Operator : JC/MD
Sample : M4817-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A6879DL

Quant Time: Nov 26 10:36:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

