

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
 Data File : VX025331.D
 Acq On : 24 Nov 2021 16:59
 Operator : JC/MD
 Sample : M4803-06 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 261-262

Quant Time: Nov 26 10:36:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206294	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	69941	50.481	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.960%
35) Dibromofluoromethane	5.391	113	63177	49.913	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	238223	50.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.085	95	84667	47.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.060%
Target Compounds						
16) Acetone	2.380	43	3054	3.967	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
Data File : VX025331.D
Acq On : 24 Nov 2021 16:59
Operator : JC/MD
Sample : M4803-06 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
261-262

Quant Time: Nov 26 10:36:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

