

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX112421\
 Data File : VX025315.D
 Acq On : 24 Nov 2021 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 26 10:30:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	50.000	56.173	-12.3	89	0.00
3 P	Chloromethane	50.000	45.618	8.8	79	0.00
4 C	Vinyl Chloride	50.000	47.217	5.6#	81	0.00
5 T	Bromomethane	50.000	48.464	3.1	86	0.00
6 T	Chloroethane	50.000	44.774	10.5	84	0.00
7 T	Trichlorofluoromethane	50.000	47.706	4.6	83	0.00
8 T	Diethyl Ether	50.000	46.396	7.2	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.021	-2.0	92	0.00
10 T	Methyl Iodide	50.000	50.745	-1.5	83	0.00
11 T	Tert butyl alcohol	250.000	206.064	17.6	72	0.00
12 CM	1,1-Dichloroethene	50.000	48.228	3.5#	86	0.00
13 T	Acrolein	250.000	242.119	3.2	84	0.00
14 T	Allyl chloride	50.000	45.355	9.3	85	0.00
15 T	Acrylonitrile	250.000	227.161	9.1	81	0.00
16 T	Acetone	250.000	239.928	4.0	86	0.00
17 T	Carbon Disulfide	50.000	42.407	15.2	78	0.00
18 T	Methyl Acetate	50.000	43.566	12.9	81	0.00
19 T	Methyl tert-butyl Ether	50.000	47.657	4.7	87	0.00
20 T	Methylene Chloride	50.000	45.218	9.6	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.898	4.2	87	0.00
22 T	Diisopropyl ether	50.000	48.533	2.9	89	0.00
23 T	Vinyl Acetate	250.000	237.641	4.9	85	0.00
24 P	1,1-Dichloroethane	50.000	48.969	2.1	89	0.00
25 T	2-Butanone	250.000	231.511	7.4	82	0.00
26 T	2,2-Dichloropropane	50.000	48.946	2.1	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.663	2.7	88	0.00
28 T	Bromochloromethane	50.000	51.226	-2.5	91	-0.01
29 T	Tetrahydrofuran	250.000	221.516	11.4	80	0.00
30 C	Chloroform	50.000	49.649	0.7#	90	0.00
31 T	Cyclohexane	50.000	47.781	4.4	87	0.00
32 T	1,1,1-Trichloroethane	50.000	48.999	2.0	89	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.521	3.0	88	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	88	0.00
35 S	Dibromofluoromethane	50.000	49.127	1.7	88	0.00
36 T	1,1-Dichloropropene	50.000	49.138	1.7	88	0.00
37 T	Ethyl Acetate	50.000	45.743	8.5	82	0.00
38 T	Carbon Tetrachloride	50.000	49.331	1.3	87	0.00
39 T	Methylcyclohexane	50.000	48.801	2.4	85	0.00
40 TM	Benzene	50.000	48.923	2.2	88	0.00
41 T	Methacrylonitrile	50.000	44.826	10.3	83	-0.01
42 TM	1,2-Dichloroethane	50.000	50.365	-0.7	89	0.00
43 T	Isopropyl Acetate	50.000	46.584	6.8	83	0.00
44 TM	Trichloroethene	50.000	49.893	0.2	88	0.00
45 C	1,2-Dichloropropane	50.000	49.520	1.0#	87	0.00
46 T	Dibromomethane	50.000	49.525	1.0	88	0.00
47 T	Bromodichloromethane	50.000	48.683	2.6	86	0.00
48 T	Methyl methacrylate	50.000	45.971	8.1	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
 Data File : VX025315.D
 Acq On : 24 Nov 2021 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 26 10:30:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	843.065	15.7	72	0.00
50 S	Toluene-d8	50.000	48.834	2.3	86	0.00
51 T	4-Methyl-2-Pentanone	250.000	231.356	7.5	82	0.00
52 CM	Toluene	50.000	49.070	1.9#	87	0.00
53 T	t-1,3-Dichloropropene	50.000	48.940	2.1	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.440	3.1	85	0.00
55 T	1,1,2-Trichloroethane	50.000	50.150	-0.3	88	0.00
56 T	Ethyl methacrylate	50.000	47.911	4.2	85	0.00
57 T	1,3-Dichloropropane	50.000	49.245	1.5	88	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.075	-6.0	96	0.00
59 T	2-Hexanone	250.000	232.810	6.9	81	0.00
60 T	Dibromochloromethane	50.000	47.939	4.1	85	0.00
61 T	1,2-Dibromoethane	50.000	48.929	2.1	87	0.00
62 S	4-Bromofluorobenzene	50.000	49.179	1.6	89	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	89	0.00
64 T	Tetrachloroethene	50.000	50.989	-2.0	91	0.00
65 PM	Chlorobenzene	50.000	49.767	0.5	89	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.270	3.5	86	0.00
67 C	Ethyl Benzene	50.000	49.808	0.4#	88	0.00
68 T	m/p-Xylenes	100.000	97.515	2.5	87	0.00
69 T	o-Xylene	50.000	49.539	0.9	88	0.00
70 T	Styrene	50.000	49.574	0.9	88	0.00
71 P	Bromoform	50.000	45.668	8.7	80	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	87	0.00
73 T	Isopropylbenzene	50.000	49.738	0.5	89	0.00
74 T	N-ethyl acetate	50.000	46.771	6.5	83	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.161	5.7	85	0.00
76 T	1,2,3-Trichloropropane	50.000	48.712	2.6	85	0.00
77 T	Bromobenzene	50.000	49.431	1.1	87	0.00
78 T	n-propylbenzene	50.000	50.895	-1.8	88	0.00
79 T	2-Chlorotoluene	50.000	49.177	1.6	89	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.464	1.1	89	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.862	10.3	78	0.00
82 T	4-Chlorotoluene	50.000	49.054	1.9	88	0.00
83 T	tert-Butylbenzene	50.000	49.096	1.8	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.204	-0.4	88	0.00
85 T	sec-Butylbenzene	50.000	50.942	-1.9	89	0.00
86 T	p-Isopropyltoluene	50.000	50.303	-0.6	88	0.00
87 T	1,3-Dichlorobenzene	50.000	49.768	0.5	87	0.00
88 T	1,4-Dichlorobenzene	50.000	49.332	1.3	89	0.00
89 T	n-Butylbenzene	50.000	50.579	-1.2	89	0.00
90 T	Hexachloroethane	50.000	45.920	8.2	81	0.00
91 T	1,2-Dichlorobenzene	50.000	49.273	1.5	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.045	5.9	79	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.849	-9.7	88	0.00
94 T	Hexachlorobutadiene	50.000	48.633	2.7	88	0.00
95 T	Naphthalene	50.000	50.478	-1.0	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112421\
Data File : VX025315.D
Acq On : 24 Nov 2021 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 26 10:30:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.848	-9.7	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6